

## CERTIFICATE OF QUALITY

|                         |   |
|-------------------------|---|
| ITEM name and code      | #3125 Handling Pipette - Handling Pipette |
| LOT NUMBER              | 20250402                                  |
| EXPIRY DATE             | 2 April 2029                              |
| STERILIZATION DATE      | 5 April 2025                              |
| STERILIZATION TECHNIQUE | Gamma Irradiation                         |
| STERILIZATION DOSE      | 22,5 - 45 kGy                             |

## QUALITY CONTROL TESTING PROCEDURES

| TEST  | SPECIFICATION  | CONCLUSION |
|---|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST  | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Mirjam Smeets  
Gynétics Medical Products N.V.

Signature date:  
2 June 2025