

## CERTIFICATE OF QUALITY

ITEM name and code	#4219 Emtrac A - Embryo Transfer Cathether
LOT NUMBER	20250513
EXPIRY DATE	13 May 2029
STERILIZATION DATE	16 May 2025
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGy

## QUALITY CONTROL TESTING PROCEDURES

TEST	SPECIFICATION	CONCLUSION
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	SPECIFICATIONS	CONCLUSION
Mouse Embryo Assay (MEA)	> 80%	Pass *
<i>expressed as day 5 expanded blastocyst stage</i>		

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Mirjam Smeets  
Gynétics Medical Products N.V.

Signature date:  
9 July 2025