CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Long - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20250923
EXPIRY DATE	23 September 2029
STERILIZATION DATE	27 October 2025
STERILIZATION TECHNIQUE	Ethylene Oxide
RESULTS OF BI	No growth of organisms

QUALITY CONTROL TESTING PROCEDURES

TEST	**SPECIFICATION**	**CONCLUSION**
Endotoxin Assay (LAL)	<0.5 EU/device	Pass
TEST	**SPECIFICATIONS**	**CONCLUSION**
Mouse Embryo Assay	> 80%	Pass
MEA Control	> 80%	Pass

Leen van Brandt

Gynétics Medical Products N.V.

Signature date:

28 November 2025