CERTIFICATE OF QUALITY

ITEM name and code	#4225 Smooze Long - Intra-Uterine Insemination Catheter/Cannula Soft
LOT NUMBER	20250924
EXPIRY DATE	24 September 2029
STERILIZATION DATE	30 September 2025
STERILIZATION TECHNIQUE	Ethylene Oxide
STERILIZATION DOSE	22,5 - 45 kGy

QUALITY CONTROL TESTING PROCEDURES

TEST	**SPECIFICATION**	**CONCLUSION**
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	**SPECIFICATIONS**	**CONCLUSION**
Mouse Embryo Assay (MEA) *expressed as day 5 expanded blastocyst stage*	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Leen van Brandt

Gynétics Medical Products N.V.

Signature date:

24 November 2025