

## CERTIFICATE OF QUALITY

**ITEM name and code** #4164 Probet - Endometrial Suction Curette

**LOT NUMBER** 20260324

**EXPIRY DATE** 24 March 2030

**STERILIZATION DATE** 30 March 2026

**STERILIZATION TECHNIQUE** Ethylene Oxide

Biological Indicator testing: no growth of organisms

## QUALITY CONTROL TESTING PROCEDURES

| TEST                                                | SPECIFICATION  | CONCLUSION |
|-----------------------------------------------------|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST                                                | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Leen van Brandt  
Gynétics Medical Products N.V.

Signature date:  
22 June 2026