

## CERTIFICATE OF QUALITY

|                         |                                        |
|-------------------------|----------------------------------------|
| ITEM name and code      | #4551 FAS - Follicle Aspiration Needle |
| LOT NUMBER              | 20260624                               |
| EXPIRY DATE             | 24 June 2030                           |
| STERILIZATION DATE      | 30 June 2026                           |
| STERILIZATION TECHNIQUE | Gamma Irradiation                      |
| STERILIZATION DOSE      | 22,5 - 45 kGy                          |

## QUALITY CONTROL TESTING PROCEDURES

| TEST                                                | SPECIFICATION  | CONCLUSION |
|-----------------------------------------------------|----------------|------------|
| Endotoxin Assay (LAL)                               | <1.0 EU/device | Pass *     |
| TEST                                                | SPECIFICATIONS | CONCLUSION |
| Mouse Embryo Assay (MEA)                            | > 80%          | Pass *     |
| <i>expressed as day 5 expanded blastocyst stage</i> |                |            |

\* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan



Leen van Brandt  
Gynétics Medical Products N.V.

Signature date:  
24 August 2026