CERTIFICATE OF QUALITY

ITEM name and code	#4551 FAS - Follicle Aspiration Needle
LOT NUMBER	20260818
EXPIRY DATE	18 August 2030
STERILIZATION DATE	22 August 2026
STERILIZATION TECHNIQUE	Gamma Irradiation
STERILIZATION DOSE	22,5 - 45 kGy

QUALITY CONTROL TESTING PROCEDURES

TEST	**SPECIFICATION**	**CONCLUSION**
Endotoxin Assay (LAL)	<1.0 EU/device	Pass *
TEST	**SPECIFICATIONS**	**CONCLUSION**
Mouse Embryo Assay (MEA) *expressed as day 5 expanded blastocyst stage*	> 80%	Pass *

* LAL and MEA results are conform the internal quality control procedure in which testing is conducted based on a representative sampling plan

Leen van Brandt

Gynétics Medical Products N.V.

Signature date:

9 September 2026