

CERTIFICATE OF ANALYSIS

PRODUCT DESCRIPTION:

EZ-Squeeze™

LOT NUMBER(S):

209147, 209148, 209149, 209173, 209174, 209175, 209176.

209177, 209178

EXPIRY DATE:

2024-09-30

INDICATIONS FOR USE:

The EZ-Squeeze pipettes are for denudation, i.e. removing the cumulus from an occute prior to Intracytoplasmic sperm

cumulus from an oocyte prior to Intracytoplasmic sperm injection (ICSI) and in vitro fertilization (IVF), and for handling

gametes, embryos and biopsied cells (polar bodies,

blastomeres and trophectoderm) during assisted reproductive

techniques (ART)..

QUALITY ASSURANCE:

Sterilised by gamma irradiation using dose between 25 and

35kGy to a SAL of 10^{-6}

Assay Test. Pass criteria:

1-cell MEA ≥ 80% hatched blastocysts at 120 hrs 100% PASS

Limulus Amoebocyte Lysate Test (LAL). Pass criteria

<20Eu/device. PASS

STORAGE CONDITIONS:

Temperature 10-30°C

Humidity 20-80% RH

Keep away from direct sunlight (UV)

Quality Assurance:

HA

Date: 22 November 2019

Caution: US Federal law restricts this device to sale by or on the order of a physician.

Research Instruments Ltd, Bickland Industrial Park, Falmouth, Cornwall, TR11 4TA, UK
Tel: +44 1326 372 753 Fax: +44 1326 378 783

customer.service@origio.com www.origio.com



CooperSurgical Fertility Companies