

PRODUCT: Vitrification Media

LOT#: N1201
 EXPIRY DATE: 2025-12-02
 Year / Month / Date

REF: 91101
 CODE: VT601



(01) 1 4582231 46207 7 (17) 251202 (10) N1201

APPROVED BY: 
 Kazumi Shibata
 Quality Control
 2025-01-14

CERTIFICATE OF ANALYSIS

Indication for Use: Vitrification Media is used for vitrification and storage of oocytes and/or embryos.

Storage: Store between 2-8 °C.

Caution: Federal law restricts this device to sale by or on the order of a physician or practitioner trained in its use.

The product has been manufactured by Kitazato Corporation in conformity with requirements of ISO13485 Quality management system. The quality of products was tested and evaluated by quality control procedures.

ASSAY	SPECIFICATION	RESULT			
		BS	ES	VS1	VS2
Color	Clear, Particle Free/ Rose Pink	Pass	Pass	Pass	Pass
pH	7.20-7.60	7.39	7.41	7.42	7.43
Osmolality ¹	BS:240-300	284	2699	5595	5595
	ES:2300-2800 VS:4900-6000				
Endotoxin ²	< 0.25 EU/mL	<0.007	<0.007	<0.007	<0.007
Sterility ³ MEA(One Cell) ⁴	No Microbial Growth	Pass	Pass	Pass	Pass
	-Control		≥ 80%	86.67%	
	-This Lot		≥ 80%	85.71%	
Gentamicin ⁵	20-50µg/mL	Pass	Pass	Pass	Pass

¹ Osmotic pressure is measured either by osmolality; with freezing point in mOsm/L or by osmolality; with vapor pressure osmometry in mmol/kg.

² Bacteria Endotoxins Test, Photometric Techniques in accordance with the USP, EP and JP

³ Sterility Tests, Direct Inoculation in accordance with the USP, EP and JP

⁴ One cell mouse embryo testing is performed using fresh mouse embryos. Test results indicate the percentage of mouse embryos developing to fully expanded blastocysts after 96 hours in culture medium.

⁵ Gentamicin quantitative assay is conducted. The acceptance criteria required is less than 50µg/mL.