



Certificate of Analysis

Product: SpermCryo™ All-round

Batch number: CRY001
Expiration Date: 30-09-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,47
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,1 EU/ml
Sperm Survival: (4 hours exposure)	≥ 80%	93%

Approved on: 23-03-2020
(DD/MM/YYYY)
By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.