



Certificate of Analysis

Product: SpermCryo™ All-round

Batch number: CRY106
Expiration Date: 31-03-2023
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,45
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi: (Sterile by filtration over 0,2µm filter)	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,1 EU/ml
HSA concentration:	3,2 – 4,8 g/liter	3,991
Sperm Survival: (4 hours exposure)	≥ 80%	92%

Approved on: 05-10-2021
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp or background.

Quality Control Department, Gynotec B.V.