



Certificate of Analysis

Product: SpermCryo[®] All-round

Batch number: CRY2603
Expiration Date: 31-07-2027
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,50
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,10
HSA concentration:	3,20 – 4,80 g/liter	3,90
HSSA (<i>motility compared with control after 4h exposure to test medium</i>):	≥ 80%	97%

Approved on: 17-02-2026
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', with the name printed below it.