



## Certificate of Analysis

### Product: SpermCryo<sup>®</sup> All-round

Batch number: CRY2605  
Expiration Date: 30-09-2027  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,40
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,10
HSA concentration:	3,20 – 4,80 g/liter	3,70
HSSA ( <i>motility compared with control after 4h exposure to test medium</i> ):	≥ 80%	98%

---

Approved on: 19-03-2026  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, consisting of a stylized 'M' and 'B' followed by the name 'M. Berbers' written in a cursive script.