



## Certificate of Analysis

**Product: SpermCryo™ All-round**

Batch number: CRY706  
Expiration Date: 31-03-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,44
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,159 EU/ml
Sperm Survival: (4 hours exposure)	≥ 80%	99%

---

Approved on: 02-10-2017  
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.