



Certificate of Analysis

Product: SpermCryo™ All-round

Batch number: CRY804
Expiration Date: 31-01-2020
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,54
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,25 EU/ml	< 0,1 EU/ml
Sperm Survival: (4 hours exposure)	≥ 80%	98%

Approved on: 26-07-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.