



## Certificate of Analysis

**Product: SpermFilter® 100%**

Batch number: FLT100-012  
Expiration Date: 31-12-2021  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,30
Osmolality:	300 – 330 mOsm/kg	313
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,100 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1197
Sperm Survival:	≥ 80%	98%

---

Approved on: 16-07-2020  
(DD/MM/YYYY)

By: J. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'J. Berbers', is written over a light blue horizontal line.

*Quality Control Department, Gynotec B.V.*