



## Certificate of Analysis

**Product: SpermFilter® 100%**

Batch number: FLT100-015  
Expiration Date: 28-02-2022  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,24
Osmolality:	300 – 330 mOsm/kg	306
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,100 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1183
Sperm Survival:	≥ 80%	98%

---

Approved on: 01-09-2020  
(DD/MM/YYYY)

By: J. Berbers

Signature:

Quality Control Department, Gynotec B.V.