



## Certificate of Analysis

**Product: SpermFilter® 100%**

Batch number: FLT100-113  
Expiration Date: 30-11-2022  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,34
Osmolality:	300 – 330 mOsm/kg	326
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1205
Sperm Survival:	≥ 80%	100%

---

Approved on: 03-06-2021  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, circular stamp or watermark.

*Quality Control Department, Gynotec B.V.*