



Certificate of Analysis

Product: SpermFilter[®] 100%

Batch number: FLT100-114
Expiration Date: 31-12-2022
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,35
Osmolality:	300 – 330 mOsm/kg	321
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1212
Sperm Survival:	≥ 80%	98%

Approved on: 29-06-2021
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, circular stamp.

Quality Control Department, Gynotec B.V.