



Certificate of Analysis

Product: SpermFilter[®] Stock Solution 100%

Batch number: FLT100-2439
Expiration Date: 30-04-2026
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,30
Osmolality:	300 – 330 mOsm/kg	318
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,50 EU/ml	< 0,15
Density:	1,1150 – 1,1250 g/ml	1,1199
HSSA: (motility compared with control after 4h exposure to test medium)	≥ 80%	99%
HSSA: (motility compared with control after 24h exposure to test medium)	≥ 75%	100%

Approved on: 12-11-2024
(DD/MM/YYYY)
By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, light blue rectangular stamp.