



## Certificate of Analysis

### Product: SpermFilter<sup>®</sup> Stock Solution 100%

Batch number: FLT100-2503  
Expiration Date: 31-07-2026  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,29
Osmolality:	300 – 330 mOsm/kg	312
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,50 EU/ml	< 0,15
Density:	1,1150 – 1,1250 g/ml	1,1204
HSSA: (motility compared with control after 4h exposure to test medium)	≥ 80%	100%
HSSA: (motility compared with control after 24h exposure to test medium)	≥ 75%	95%

---

Approved on: 12-02-2025  
(DD/MM/YYYY)  
By: M. Berbers

Signature: