



Certificate of Analysis

Product: SpermFilter® Stock Solution 100%

Batch number: FLT100-2507

Expiration Date: 31-08-2026
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,30
Osmolality:	300 – 330 mOsm/kg	321
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,50 EU/ml	< 0,15
Density:	1,1150 – 1,1250 g/ml	1,1193
HSSA:	≥ 80%	100%
(motility compared with control after 4h exposure to test medium)		
HSSA:	≥ 75%	91%
(motility compared with control after 24h exposure to test medium)		

Approved on: 10-03-2025

(DD/MM/YYYY)

By: M. Berbers

Signature: