



Certificate of Analysis

Product: SpermFilter® 100%

Batch number: FLT100-323
Expiration Date: 31-01-2025
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,41
Osmolality:	300 – 330 mOsm/kg	316
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,50 EU/ml	< 0,10
Density:	1,1150 – 1,1250 g/ml	1,1208
Viscosity in cP (25°C):	-	1,71
Sperm Survival: (4 hours exposure)	≥ 80%	97%
Sperm Survival: (24 hours exposure)	≥ 75%	94%

Approved on: 24-07-2023
(DD/MM/YYYY)
By: M. Berbers

Signature: