



Certificate of Analysis

Product: SpermFilter[®] 100%

Batch number: FLT100-722
Expiration Date: 31-05-2019
(DD/MM/YYYY)

| Analysis | Limits | Results |
|--------------------------|----------------------|-------------|
| pH: | 7,20 – 7,60 | 7,55 |
| Osmolality: | 300 – 330 mOsm/kg | 309 |
| Sterility Test Bacteria: | NEGATIVE | PASSED |
| Sterility Test Fungi: | NEGATIVE | PASSED |
| Endotoxin: | < 0,5 EU/ml | < 0,1 EU/ml |
| Density: | 1,1150 – 1,1250 g/ml | 1,1193 |
| Sperm Survival: | ≥ 80% | 97% |

Approved on: 13-12-2017
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.