



## Certificate of Analysis

**Product: SpermFilter<sup>®</sup> 100%**

Batch number: FLT100-804  
Expiration Date: 31-08-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,52
Osmolality:	300 – 330 mOsm/kg	319
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1192
Sperm Survival:	≥ 80%	95%

---

Approved on: 23-02-2018  
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.