



## Certificate of Analysis

**Product: SpermFilter® 100%**

Batch number:

FLT100-812

Expiration Date:  
(DD/MM/YYYY)

30-11-2019

| Analysis                 | Limits               | Results     |
|--------------------------|----------------------|-------------|
| pH:                      | 7,20 – 7,60          | 7,31        |
| Osmolality:              | 300 – 330 mOsm/kg    | 305         |
| Sterility Test Bacteria: | NEGATIVE             | PASSED      |
| Sterility Test Fungi:    | NEGATIVE             | PASSED      |
| Endotoxin:               | < 0,5 EU/ml          | 0,174 EU/ml |
| Density:                 | 1,1150 – 1,1250 g/ml | 1,1197      |
| Sperm Survival:          | ≥ 80%                | 98%         |

Approved on:  
(DD/MM/YYYY)

25-05-2018

By:

J. Berbers

Signature:

Quality Control Department, Gynotec B. V.