Gynotec

# Certificate of Analysis

## Product: SpermFilter® 100%

| | |
|---|---|
| Batch number: | FLT100-829 |
| Expiration Date: (DD/MM/YYYY) | 31-05-2020 |

| Analysis | Limits | Results |
|---|---|---|
| pH: | 7,20 – 7,60 | 7,34 |
| Osmolality: | 300 – 330 mOsm/kg | 309 |
| Sterility Test Bacteria: | NEGATIVE | PASSED |
| Sterility Test Fungi: | NEGATIVE | PASSED |
| Endotoxin: | < 0,5 EU/ml | <0,1 EU/ml |
| Density: | 1,1150 – 1,1250 g/ml | 1,1233 |
| Sperm Survival: | ≥ 80% | 93% |

---

| | |
|---|---|
| Approved on: (DD/MM/YYYY) | 12-12-2018 |
| By: | M. Berbers |
| Signature: | |

*Quality Control Department, Gynotec B.V.*

Ref. nr.: F050/3