Gynotec

Certificate of Analysis

Product: SpermFilter® 100%

Batch number:	FLT100-910
Expiration Date: (DD/MM/YYYY)	30-09-2020

Analysis	Limits	Results
pH:	7,20 – 7,60	7,34
Osmolality:	300 – 330 mOsm/kg	320
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,167 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1198
Sperm Survival:	≥ 80%	98%

Approved on: (DD/MM/YYYY)	10-04-2019
By:	M. Berbers
Signature:	

Quality Control Department, Gynotec B.V.

Ref. nr.: F050/3