



Certificate of Analysis

Product: SpermFilter® 100%

Batch number: FLT100-920
Expiration Date: 28-02-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,35
Osmolality:	300 – 330 mOsm/kg	318
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,100 EU/ml
Density:	1,1150 – 1,1250 g/ml	1,1205
Sperm Survival:	≥ 80%	97%

Approved on: 09-09-2019
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.