



## Certificate of Analysis

**Product: SpermFilter® 100%**

Batch number: FLT100-930  
Expiration Date: 31-05-2021  
(DD/MM/YYYY)

| Analysis                 | Limits               | Results     |
|--------------------------|----------------------|-------------|
| pH:                      | 7,20 – 7,60          | 7,32        |
| Osmolality:              | 300 – 330 mOsm/kg    | 311         |
| Sterility Test Bacteria: | NEGATIVE             | PASSED      |
| Sterility Test Fungi:    | NEGATIVE             | PASSED      |
| Endotoxin:               | < 0,5 EU/ml          | 0,100 EU/ml |
| Density:                 | 1,1150 – 1,1250 g/ml | 1,1198      |
| Sperm Survival:          | ≥ 80%                | 99%         |

---

Approved on: 02-12-2019  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, circular stamp.

Quality Control Department, Gynotec B.V.