



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number: FLT45-009
Expiration Date: 30-09-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,28
Osmolality:	310 – 340 mOsm/kg	329
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	98%

Approved on: 31-03-2020
(DD/MM/YYYY)
By: J. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'J. Berbers', is written over a light blue horizontal line.

Quality Control Department, Gynotec B.V.