Gynotec

Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number:	FLT45-037
Expiration Date: (DD/MM/YYYY)	31-05-2022

Analysis	Limits	Results
pH:	7,20 – 7,60	7,36
Osmolality:	310 – 340 mOsm/kg	318
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	99%

Approved on: (DD/MM/YYYY)	01-12-2020
By:	M. Berbers
Signature:	

Quality Control Department, Gynotec B.V.