



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number: FLT45-127
Expiration Date: 28-02-2023
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,34
Osmolality:	310 – 340 mOsm/kg	334
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
<i>(Sterile by filtration over 0,2µm filter)</i>		
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	97%
<i>(4 hours exposure)</i>		
Sperm Survival:	≥ 75%	95%
<i>(24 hours exposure)</i>		

Approved on: 27-08-2021
(DD/MM/YYYY)
By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.