



Certificate of Analysis

Product: SpermFilter[®] Gradient 45%

Batch number: FLT45-206
Expiration Date: 31-08-2023
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,41
Osmolality:	310 – 340 mOsm/kg	325
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
<i>(Sterile by filtration over 0,2µm filter)</i>		
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	99%
<i>(4 hours exposure)</i>		
Sperm Survival:	≥ 75%	97%
<i>(24 hours exposure)</i>		

Approved on: 09-03-2022
(DD/MM/YYYY)
By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.