



## Certificate of Analysis

### Product: SpermFilter<sup>®</sup> Gradient 45%

Batch number: FLT45-2543  
Expiration Date: 28-02-2027  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,29
Osmolality:	280 – 310 mOsm/kg	301
Sterility Test (According to current Ph. Eur. 2.6.1 / USP <71>):	No Growth	PASSED
Endotoxin:	< 0,50 EU/ml	< 0,15
Density:	1,0500 – 1,0700 g/ml	1,0579
HSA concentration:	1,80 – 2,60 g/liter	2,00
Gentamicin concentration:	5 – 15 mg/liter	13,00
HSSA (motility compared with control after 4h exposure to test medium):	≥ 80%	100%
HSSA (motility compared with control after 24h exposure to test medium):	≥ 75%	98%

---

Approved on: 09-09-2025  
(DD/MM/YYYY)  
By: M. Berbers

Signature: