



## Certificate of Analysis

**Product: SpermFilter<sup>®</sup> Gradient 45%**

Batch number: FLT45-735  
Expiration Date: 31-03-2019  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,40
Osmolality:	310 – 340 mOsm/kg	311
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Sperm Survival:	≥ 80%	98%

---

Approved on: 17-10-2017  
(DD/MM/YYYY)

By: M. Berbers

Signature:

*Quality Control Department, Gynotec B.V.*