



Certificate of Analysis

Product: SpermFilter[®] Gradient 45%

Batch number: FLT45-811
Expiration Date: 30-09-2019
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,37
Osmolality:	310 – 340 mOsm/kg	324
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Sperm Survival:	≥ 80%	100%

Approved on: 21-03-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.