



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number:

FLT45-820

Expiration Date:
(DD/MM/YYYY)

30-11-2019

Analysis	Limits	Results
pH:	7,20 – 7,60	7,41
Osmolality:	310 – 340 mOsm/kg	320
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Sperm Survival:	≥ 80%	97%

Approved on:
(DD/MM/YYYY)

24-05-2018

By:

J. Berbers

Signature:

Quality Control Department, Gynotec B. V.