



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number: FLT45-827
Expiration Date: 31-01-2020
(DD/MM/YYYY)

| Analysis | Limits | Results |
|--------------------------|-------------------|-------------|
| pH: | 7,20 – 7,60 | 7,38 |
| Osmolality: | 310 – 340 mOsm/kg | 321 |
| Sterility Test Bacteria: | NEGATIVE | PASSED |
| Sterility Test Fungi: | NEGATIVE | PASSED |
| Endotoxin: | < 0,5 EU/ml | < 0,1 EU/ml |
| Sperm Survival: | ≥ 80% | 99% |

Approved on: 10-08-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.