



Certificate of Analysis

Product: SpermFilter[®] Gradient 45%

Batch number: FLT45-834
Expiration Date: 30-04-2020
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,34
Osmolality:	310 – 340 mOsm/kg	319
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	0,215 EU/ml
Sperm Survival:	≥ 80%	100%

Approved on: 19-10-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.