



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number: FLT45-837
Expiration Date: 30-04-2020
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,36
Osmolality:	310 – 340 mOsm/kg	322
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Sperm Survival:	≥ 80%	100%

Approved on: 19-11-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to be 'M. Berbers', written over a horizontal line.

Quality Control Department, Gynotec B.V.