



Certificate of Analysis

Product: SpermFilter® Gradient 45%

Batch number: FLT45-934
Expiration Date: 30-04-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,34
Osmolality:	310 – 340 mOsm/kg	325
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	97%

Approved on: 28-10-2019
(DD/MM/YYYY)
By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp.

Quality Control Department, Gynotec B.V.