



## Certificate of Analysis

### Product: SpermFilter® Gradient 45%

Batch number: FLT45-942  
Expiration Date: 30-06-2021  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,31
Osmolality:	310 – 340 mOsm/kg	325
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,100 EU/ml
Sperm Survival:	≥ 80%	98%

---

Approved on: 19-12-2019  
(DD/MM/YYYY)  
By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, rectangular stamp.

Quality Control Department, Gynotec B.V.