



## Certificate of Analysis

### Product: SpermFilter® Gradient 80%

Batch number: FLT80-135  
Expiration Date: 30-04-2023  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,35
Osmolality:	310 – 340 mOsm/kg	325
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
<i>(Sterile by filtration over 0,2µm filter)</i>		
Endotoxin:	< 0,5 EU/ml	0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,1020
Viscosity in cP (25°)	< 1,65 cP	1,63
Sperm Survival:	≥ 80%	99%
<i>(4 hours exposure)</i>		

---

Approved on: 08-11-2021  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', is written over a faint, illegible stamp or background.

Quality Control Department, Gynotec B.V.