



Certificate of Analysis

Product: SpermFilter[®] Gradient 80%

Batch number: FLT80-740
Expiration Date: 30-04-2019
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,49
Osmolality:	310 – 330 mOsm/kg	310
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,0973
Sperm Survival:	≥ 80%	100%

Approved on: 13-11-2017
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.