



## Certificate of Analysis

**Product: SpermFilter<sup>®</sup> Gradient 80%**

Batch number: FLT80-830  
Expiration Date: 31-03-2020  
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,47
Osmolality:	310 – 330 mOsm/kg	321
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,5 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,0984
Sperm Survival:	≥ 80%	98%

---

Approved on: 26-09-2018  
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to be 'M. Berbers', written over a horizontal line.

*Quality Control Department, Gynotec B.V.*