



Certificate of Analysis

Product: SpermFilter[®] Gradient 80%

Batch number: FLT80-841
Expiration Date: 31-05-2020
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,36
Osmolality:	310 – 330 mOsm/kg	316
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,1001
Sperm Survival:	≥ 80%	96%

Approved on: 11-12-2018
(DD/MM/YYYY)

By: M. Berbers

Signature:

Quality Control Department, Gynotec B.V.