Gynotec

Certificate of Analysis

Product: SpermFilter® Gradient 80%

Batch number:	FLT80-843
Expiration Date: (DD/MM/YYYY)	30-06-2020

Analysis	Limits	Results
pH:	7,20 – 7,60	7,37
Osmolality:	310 – 330 mOsm/kg	318
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,1 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,1008
Sperm Survival:	≥ 80%	97%

Approved on: (DD/MM/YYYY)	27-12-2018
By:	M. Berbers
Signature:	

Quality Control Department, Gynotec B.V.

Ref. nr.: F052/3