



Certificate of Analysis

Product: SpermFilter® Gradient 80%

Batch number: FLT80-925
Expiration Date: 31-01-2021
(DD/MM/YYYY)

Analysis	Limits	Results
pH:	7,20 – 7,60	7,37
Osmolality:	310 – 330 mOsm/kg	323
Sterility Test Bacteria:	NEGATIVE	PASSED
Sterility Test Fungi:	NEGATIVE	PASSED
Endotoxin:	< 0,5 EU/ml	< 0,311 EU/ml
Density:	1,0970 – 1,1070 g/ml	1,0985
Sperm Survival:	≥ 80%	95%

Approved on: 30-07-2019
(DD/MM/YYYY)

By: M. Berbers

Signature:

A handwritten signature in blue ink, appearing to read 'M. Berbers', with the name printed below it.

Quality Control Department, Gynotec B.V.