

PRODUCT: Thawing Media

LOT#: H0220

EXPIRY DATE: 2018-06-08

Year / Month / Date

REF: 91182

CODE: VT802



(01) 1 4582231 46077 6 (17) 180608 (10) H0220

*Kazumi Shibata*  
 APPROVED BY: Kazumi Shibata  
 Quality Assurance  
 2018-03-08

**CERTIFICATE OF ANALYSIS**

**Indication for Use:** Thawing Media is used for thawing and recovery of vitrified oocytes and/or embryos.

**Storage:** Store between 2-8 °C.

**Caution:** Federal law restricts this device to sale by or on the order of a physician or practitioner trained in its use.

We certify that products manufactured by Kitazato Corporation have been approved in accordance with current Good Manufacturing Practice (cGMP). The quality of products was tested and evaluated by quality control procedures.

ASSAY	SPECIFICATION	RESULT			
		TS1	TS2	DS	WS
Color	Clear, Particle Free/ Rose Pink	Pass	Pass	Pass	Pass
pH	7.20-7.60	7.39	7.41	7.40	7.40
Osmolarity <sup>1</sup>	TS:1600-2000				
	DS:830-1020	1902	1904	940	286
	WS:240-300				
Endotoxin <sup>2</sup>	< 0.25 EU/mL	<0.006	<0.006	<0.006	<0.006
Sterility <sup>3</sup>	No Microbial Growth	Pass	Pass	Pass	Pass
MEA(One Cell) <sup>4</sup>					
-Control	≥ 80%		93.33%		
-This Lot	≥ 80%		95.24%		

<sup>1</sup> Osmotic pressure is measured either by osmolarity; with freezing point in mOsm/L or by osmolality; with vapor pressure osmometry in mmol/kg.

<sup>2</sup> Bacteria Endotoxins Test, Photometric Techniques in accordance with the USP, EP and JP

<sup>3</sup> Sterility Tests, Direct Inoculation in accordance with the USP, EP and JP

<sup>4</sup> One cell mouse embryo testing is performed using fresh mouse embryos. Test results indicate the percentage of mouse embryos developing to fully expanded blastocysts after 96 hours in culture medium.